

Valori delle ponderazioni A,B,C,D

FREQUENZA (Hz)	Ponder. A dB	Ponder. B dB	Ponder. C dB	Ponder. D dB
10	-70.4	-38.2	-14.3	-26.5
12.5	-63.4	-33.2	-11.2	-24.5
16	-56.7	-28.5	- 8.5	-22.5
20	-50.5	-24.2	- 6.2	-20.5
25	-44.7	-20.4	- 4.4	-18.5
31.5	-39.4	-17.1	- 3.0	-16.5
40	-34.6	-14.2	- 2.0	-14.5
50	-30.2	-11.6	- 1.3	-12.5
63	-26.2	- 9.3	- 0.8	-11.0
80	-22.5	- 7.4	- 0.5	- 9.0
100	-19.1	- 5.6	- 0.3	- 7.5
125	-16.1	- 4.2	- 0.2	- 6.0
160	-13.4	- 3.0	- 0.1	- 4.5
200	-10.9	- 2.0	0	- 3.0
250	- 8.6	- 1.3	0	- 2.0
315	- 6.6	- 0.8	0	- 1.0
400	- 4.8	- 0.5	0	- 0.5
500	- 3.2	- 0.3	0	0
630	- 1.9	- 0.1	0	0
800	- 0.8	0	0	0
1000	0	0	0	0
1250	+ 0.6	0	0	+ 2.0
1600	+ 1.0	0	- 0.1	+ 5.5
2000	+ 1.2	- 0.1	- 0.2	+ 8.0
2500	+ 1.3	- 0.2	- 0.3	+10.0
3150	+ 1.2	- 0.4	- 0.5	+11.0
4000	+ 1.0	- 0.7	- 0.8	+11.0
5000	+ 0.5	- 1.2	- 1.3	+10.0
6300	- 0.1	- 1.9	- 2.0	+ 8.5
8000	- 1.1	- 2.9	- 3.0	+ 6.0
10000	- 2.5	- 4.3	- 4.4	+ 3.0
12500	- 4.3	- 6.1	- 6.2	0
16000	- 6.6	- 8.4	- 8.5	- 4.0
20000	- 9.3	-11.1	-11.2	- 7.5